



**Jim Doyle, Governor**  
**Mary P. Burke, Secretary**

July 19, 2004

SIRONA DENTAL SYSTEMS, LLC  
MOHAMAD HAMDAN  
4835 SIRONA DR SUITE 100  
CHARLOTTE NC 28273

SIRONA DENTAL SYSTEMS, BMBH  
MOHAMAD HAMDAN  
FABRIKSTRASSE 31  
BENSHEIM D-64625  
GERMANY

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: SIRONA DENTAL SYSTEMS, BMBH  
Product Name: SIRONA C8+  
Model Number(s): D3417  
Product File No: 20040314

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2009.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 19990129.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Michael J. Beckwith, CIPE  
Plumbing Product Reviewer  
phone: 608-266-6742  
fax: 608-267-9566  
e-mail: mike.beckwith@wi.gov